

Certificate of Compliance

Certificate: 1070387 (LR 81532-C93)

Master Contract: 159007

Project: 1070387

Date Issued: April 27, 2000


Issued to: Industrial Scientific Corp.
1001 Oakdale Road
Oakdale, PA 15071-1500
Attention: Mr. Paul D. Kavscek

The products listed below are eligible to bear the CSA Mark shown



Issued by: Ray Lehman

Signature:



PRODUCTS

CLASS 4812 06 - SIGNAL APPLIANCES - Miscellaneous

Gas Monitor, Model AirAware 6810-0056-xxx, (xxx maybe from 1 to 3 alpha characters indicating options as described below). Input rated 24V ac 60Hz / 24Vdc 0.125A; relay contacts 30V dc, 5A, Transmitter output 4-20 ma.

C = Controller, T=Transmitter, A=Audio

Gas Monitor, Model 6810-0056-M AirAware Monitor, 24V ac, 60Hz / 24V dc, 0.125A supplied with a Certified Class 2 Power Supply Input 120V ac, 23W, output 24V ac, 0.6A.

Note: CERTIFIED FOR ELECTRICAL SHOCK AND ELECTRICAL FIRE HAZARD ONLY.

APPLICABLE REQUIREMENTS

CSA Std C22.2 No. 205-M1983 - Signal Equipment



Supplement to Certificate of Compliance

Certificate: 1070387 (LR 81532-C93)

Master Contract: 159007

Project: 1070387

Issued to: **Industrial Scientific Corp.**
1001 Oakdale Road
Oakdale, PA 15071-1500
Attention: Mr. Paul D. Kovscek

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Issued By: Ray Lehman

Signature: _____

A handwritten signature in black ink, appearing to read 'R. Lehman', written over a horizontal line.

Product Certification History

Project	Date	Description
1070387	April 27, 2000	Update to model numbering
LR 81532-C93	October 12, 1999	Original Certification